

Ziegler-Nichols Loop Tuning

The Ziegler-Nichols loop tuning protocol¹⁾ is a simple method to establishing the loop parameter values for a PI or PID control loop. This protocol is designed to optimize the loop for disturbance rejection.





Links

- Click here for the [SLICE-QT Manual](#).
- Click here for the [SLICE-QT Quick Start Guide](#).
- Click here for the [SLICE-QT API](#).
- Click here for the [SLICE-QT web page](#).

Instructions



There is the possibility for the loop to go unstable at too high gain. Use caution as gain is increased.

1. From the Home Screen, touch the Channel icon for the loop to be tuned
2. Touch Settings > PID Params
3. Choose a Set Point temperature
4. Set Proportional Gain (K_p) to 0
5. Turn off Integral and Differential portions of the loop (T_i & T_d , respectively)
6. While monitoring the error in temperature on rolling graph on the SLICE-QT display or an [external oscilloscope](#), **slowly** increase K_p until the error begins to oscillate
7. Use the  and  icons, to scale displayed response appropriately
8. Use the  icon to pause the graph
9. Note K_p where oscillation begins, $K_{p, cr}$
10. Use the Cursors enabled by touching the  icon to measure the period of oscillation, T_{cr}
11. Use the following table to set the values of K_p , T_i , and T_d

Type of Control Loop	K_p	T_i	T_d
PI	$0.45K_{p, cr}$	$0.83T_{cr}$	0
PID	$0.6K_{p, cr}$	$0.5T_{cr}$	$0.125T_{cr}$

Tab. 1: Ziegler-Nichols loop values

¹⁾
[Optimum Settings for Automatic Controllers](#), by J.G. Zieger and N. B. Nichols and [Wikipedia](#)

From:

<https://www.vescent.com/manuals/> - **Product Manuals**

Permanent link:

<https://www.vescent.com/manuals/doku.php?id=slice:z-n&rev=1554221304>

Last update: **2021/08/26 14:26**

